

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|                                                                                                                                          |                                                                                                                        |                                                                                                                                                |
|------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------|
| <br><b>29141</b><br>Accredited to<br>ISO/IEC 17025:2017 | <b>TechniVal Ltd</b><br><br><b>Issue No: 001    Issue date: 19 July 2024</b>                                           |                                                                                                                                                |
|                                                                                                                                          | <b>Block 4 Unit</b><br><b>7 Chapelhall Industrial Estate</b><br><b>Airdrie</b><br><b>Lanarkshire</b><br><b>ML6 8QH</b> | <b>Contact: Mr David Nelson</b><br><b>Tel: +44 (0)8450034247</b><br><b>E-Mail: info@technival.co.uk</b><br><b>Website: www.technival.co.uk</b> |

Calibration performed by the Organisation at the locations specified

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details                                                                                                                                     | Activity                                       | Location code |
|------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------|---------------|
| <b>Address:</b><br>Block 4 Unit<br>7 Chapelhall Industrial Estate<br>Airdrie,<br>Lanarkshire<br>ML6 8QH<br><br><b>Local contact:</b><br>David Nelson | Temperature (indicators, Dry Blocks, chambers) | P             |

#### Site activities performed away from the locations listed above:

| Location details                                                                                                                                                                        | Activity                           | Location code |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------|---------------|
| The location must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | Temperature (chambers, indicators) | S             |



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TechniVal Ltd**

**Issue No: 001    Issue date: 19 July /2024**

**Calibration performed by the Organisation at the locations specified**

**Calibration and Measurement Capability (CMC)**

| Measured Quantity<br>Instrument or Gauge                                                                                                                                                                                                | Range            | Expanded Measurement<br>Uncertainty ( <i>k</i> = 2) | Remarks                                                                                                                                                                    | Location<br>Code |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------|-----------------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------|
| <b>TEMPERATURE</b>                                                                                                                                                                                                                      |                  |                                                     |                                                                                                                                                                            |                  |
| Temperature indicators and<br>recorders, with temperature<br>sensor(s)                                                                                                                                                                  | -196 °C          | 0.067 °C                                            | LN <sup>2</sup> comparator<br>Stirred liquid baths<br><b>UKF001 Service</b><br><b>UKF008 Service (multi-<br/>probe loggers)</b>                                            | P                |
|                                                                                                                                                                                                                                         | -80 °C           | 0.033 °C                                            |                                                                                                                                                                            | P                |
|                                                                                                                                                                                                                                         | 0 to +50 °C      | 0.029 °C                                            |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | 50 °C to 250 °C  | 0.032 °C                                            |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | 300 °C           | 0.085 °C                                            |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | 0 °C to 60 °C    | 0.30 °C                                             | In Air                                                                                                                                                                     | P and S          |
|                                                                                                                                                                                                                                         | -90 °C to 0°C    | 0.10 °C                                             | Dry Well<br><b>UKF006 Service</b>                                                                                                                                          | P and S          |
|                                                                                                                                                                                                                                         | 0 to +125 °C     | 0.09 °C                                             |                                                                                                                                                                            |                  |
| 125 °C to 300 °C                                                                                                                                                                                                                        | 0.11 °C          |                                                     |                                                                                                                                                                            |                  |
| Temperature controlled<br>chambers, fridges, freezers,<br>ovens, furnaces and liquid<br>baths (inclusive of associated<br>indicators, controllers and<br>recorders, all with sensors,<br>within the specified<br>parameters and ranges) | -196 °C          | 0.085 °C                                            | Single monitoring probe.<br>Time dependent by<br>comparison<br><b>UK005 Service</b>                                                                                        | S                |
|                                                                                                                                                                                                                                         | -80 °C to 150°C  | 0.064 °C                                            |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | 150 °C to 300 °C | 0.099 °C                                            |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | -196 °C to 0 °C  | 0.40 °C                                             | Multipoint monitoring<br>probes.<br>Time dependent<br>temperature profiling,<br>also referred to as spatial<br>temperature surveying or<br>mapping<br><b>UK007 Service</b> | S                |
|                                                                                                                                                                                                                                         | 0 °C to 250 °C   | 0.37 °C                                             |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | 250 °C to 300 °C | 0.42 °C                                             |                                                                                                                                                                            |                  |
| Dry Block Calibrators                                                                                                                                                                                                                   | -90 °C to 0 °C   | 0.032 °C                                            | Method consistent with<br>Euramet CG13<br><b>UK004 Service</b>                                                                                                             | P                |
|                                                                                                                                                                                                                                         | 0 °C to +250 °C  | 0.028 °C                                            |                                                                                                                                                                            |                  |
|                                                                                                                                                                                                                                         | 250 °C to 400 °C | 0.042 °C                                            |                                                                                                                                                                            |                  |
| END                                                                                                                                                                                                                                     |                  |                                                     |                                                                                                                                                                            |                  |



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TechniVal Ltd**

**Issue No: 001    Issue date: 19 July /2024**

**Calibration performed by the Organisation at the locations specified**

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$